



**DISCOUNTS**

**A. Competition at Bench**

- 1. Failure to verbally identify each test being conducted. . . . 2 \_\_\_\_\_
- 2. Failure to verbally identify **each problem**. . . . . 5 \_\_\_\_\_
- 3. Failure to correct **each problem** . . . . . 5 \_\_\_\_\_
- 4. Continuing without correcting a predetermined problem . . .40 \_\_\_\_\_
- 5. Failure to conduct any visual examination or test on the BioPak 240-R, **each test** . . . . . 5 \_\_\_\_\_
- 6. Failure to tighten connections properly during assembly, **each connection** . . . . . 1 \_\_\_\_\_
- 7. Failure to zero the tester . . . . . 10 \_\_\_\_\_
- 8. Failure to comply with rules not covered in discount sheet, each infraction . . . . . 2 \_\_\_\_\_

**B. Visuals**

- 2. Failure to conduct a proper visual examination of the upper and lower housing/harness . . . . . 2 \_\_\_\_\_
- 3. Failure to conduct a proper visual examination of the center section assembly. . . . . 2 \_\_\_\_\_
- 4. Failure to conduct a proper visual examination of the carbon dioxide scrubber . . . . . 2 \_\_\_\_\_
- 5. Failure to conduct a proper visual examination of the center section lid assembly. . . . . 2 \_\_\_\_\_
- 6. Failure to conduct a proper visual examination of the O<sub>2</sub> cylinder . . . . . 2 \_\_\_\_\_
- 7. Failure to conduct a proper visual examination of the hoses. . . . . 2 \_\_\_\_\_
- 8. Failure to conduct a proper visual examination of the facepiece . . . . . 2 \_\_\_\_\_
- 9. Failure to have visual apparatus fully assembled (ready for use) . . . . . 5 \_\_\_\_\_

**C. Tester**

- 1. Failure to conduct the proper constant flow test . . . . . 2 \_\_\_\_\_
- 2. Failure to conduct a proper demand valve functional test . 2 \_\_\_\_\_
- 3. Failure to conduct a proper emergency bypass functional test. . . . . 2 \_\_\_\_\_

**DISCOUNTS**

- 4. Failure to conduct the proper vent valve functional test . . . . . 2 \_\_\_\_\_
- 5. Failure to conduct a proper low pressure leak test . . . . . 2 \_\_\_\_\_
- 6. Failure to conduct a proper RMS Gauge and TRIM test. . . 2 \_\_\_\_\_
- 7. Failure to conduct a proper high pressure leak test . . . . . 2 \_\_\_\_\_
- 8. Failure to conduct a proper low pressure alarm functional test . . . . . 2 \_\_\_\_\_
- 9. Failure to have test apparatus fully assembled (ready for use) . . . . . 5 \_\_\_\_\_

**General Rules**

- 1. Intentional damage to test equipment/apparatus . . . . . 100 \_\_\_\_\_

\_\_\_\_\_  
**Bench Discount**

**Written Examination**

- 1. For each incorrect answer . . . . . 1 \_\_\_\_\_

\_\_\_\_\_  
**Total Discounts**